

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	CAMPYLOBACTER BLOOD FREE AGAR (mCCD) (ISO 10272)
Code	010021
Lot number	2759r209
Production date	09/02/24
Expiration date	09/04/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	7,4 +/- 0,2	7.4 (Confirmed)
Quantity of material	20ml (010021)	Confirmed
Color & cleanliness of the material	Agar black not clear	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
1/010021	36 ± 1 °C for 48 hours	1/010021	0


**FINAL PRODUCT PERFORMANCE TEST (According to ISO 10272-2.)** Incubate: 41,5 °C ± 1°C for 44 ± 4 hours, in microaerobic conditions.

Microorganism	Expected results	Observed results
<i>Campylobacter jejuni</i> (≥500 cfu/10μl) ATCC: 29428. Lot: 08203201. (productivity)	PR ≥ 0,5 Colonies are small, gray	Colonies are small, gray. 323 cfu.
<i>Escherichia coli</i> (≥500 cfu/10μl) ATCC 25922. Lot: 00207501 (selectivity)	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth

**Release of the batch**YES  NO **DATE**

12/02/24

**Iniotaki Fofi****Quality Control Manager**