

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	CETRIMIDE AGAR
Code	010024
Lot number	2791r228
Production date	28/03/24
Expiration date	28/07/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	7,2 +/- 0,2	7,1 (Confirmed)
Quantity of material	20ml (010024)	Confirmed
Color & cleanliness of the material	Clear beige agar.	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
1/010024	36 ± 1 °C for 48 hours	1/010024	0

FINAL PRODUCT PERFORMANCE TEST Incubate: (36 ± 1) °C for (24 ± 2) h

Microorganism	Expected results	Observed results
<i>Pseudomonas aeruginosa</i> (≥500 cfu) ATCC 27853. Lot: 01007301 (productivity)	PR ≥ 0,5 Yellow-green colonies with halo	Yellow-green colonies with halo 302 cfu
<i>Escherichia coli</i> (≥500 cfu) ATCC 25922. Lot: 00207501 (selectivity)	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth


Release of the batch

YES NO

DATE

01/04/2024

Iniotaki Fofi



Quality Control Manager