

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	CHROMagar™ CANDIDA
Code	010028
Lot number	2758c108
Production date	08/02/24
Expiration date	08/05/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	6.1 +/- 0,2	6.2 (Confirmed)
Quantity of material	20ml (010028)	Confirmed
Color & cleanliness of the material	Beige agar light clear.	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
3/010028	36 ± 1 °C for 48 hours	3/010028	0


FINAL PRODUCT PERFORMANCE TEST Incubate: (36 ± 1) °C for (24 ± 2) h

Microorganism	Expected results	Observed results
<i>Candida albicans</i> (≥500 cfu/10μl) ATCC: 2091 Lot: 04205401 (productivity) Incubate: (36 ± 1) °C for (24 ± 2) h	PR ≥ 0,7 Green colonies	Green colonies 372 cfu
<i>Candida krusei</i> (≥500 cfu/10μl) NCPF 3100. Lot: 00300101 (productivity) Incubate: (36 ± 1) °C for (24 -48) h	PR ≥ 0,7 Pink colonies	Pink colonies 365 cfu
<i>Escherichia coli</i> (≥500 cfu/10μl) ATCC 25922 Lot: 00207501 (selectivity) Incubate: (36 ± 1) °C for (24 ± 2) h	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / EAOT EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.

Release of the batchYES NO **DATE**

12/02/2024

Iniotaki Fofi**Quality Control Manager**