

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	<b>YEAST EXTRACT DEXTROSE CHLORAMPHENICOL AGAR</b> CHLORAMPHENICOL AGAR (YGC AGAR) (ISO 7954)
Code	<b>010424</b>
Lot number	2791n128
Production date	28/03/24
Expiration date	28/07/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	6,6 +/- 0,2	6.7 (Confirmed)
Quantity of material	21ml (010424)	Confirmed
Color & cleanliness of the material	Beige clear	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3%)	Incubation	Number of suitable products	Number of inappropriate products
2/010424	36 ± 1 °C για 48 ώρες	3/010424	0


**FINAL PRODUCT PERFORMANCE TEST**

Μικροοργανισμός	Αναμενόμενα αποτελέσματα (Expected results)	Παρατηρούμενα αποτελέσματα (Observed results)
<i>Candida albicans</i> (≥500 cfu/10μl) e.10/24 ATCC: 10231. Lot: 04205401 (productivity) Επώαση: : 25 ± 1 °C για 3 - 5 μέρες	<i>PR</i> ≥ 0,5 Χαρακτηριστικές αποικίες σύμφωνες με το είδος	Χαρακτηριστικές αποικίες σύμφωνες με το είδος 295 cfu
<i>Escherichia coli</i> (≥500 cfu/10μl) e.10/24 ATCC: 25922. Lot: 00207501 (selectivity) Επώαση: : 25 ± 1 °C για 3 - 5 μέρες	Αναστέλλεται (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth

**Release of the batch**YES  NO **DATE**

05/04/23

**Iniotaki Fofi****Quality Control Manager**