

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	CHROMagar™ Malassezia
Code	010518
Lot number	2793e01
Production date	01/04/24
Expiration date	01/07/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	6,3 +/- 0,3	6.3 (Confirmed)
Quantity of material	20ml (010518)	Confirmed
Color & cleanliness of the material	Beige agar light clear.	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
1/010518	36 ± 1 °C for 48 hours	1/010518	0


**FINAL PRODUCT PERFORMANCE TEST**

Microorganism	Expected results	Observed results
<i>Malassezia furfur</i> (≥500 cfu/10μl) ATCC 14521. Lot: 06405831 (productivity) Incubate: (30 ± 37) °C για (72 ± 2) h	PR ≥ 0.7 Colonies pale pink and wrinkled (2-3mm)	Colonies pale pink and wrinkled (2mm) 396 cfu
<i>Candida albicans</i> (≥500 cfu/10μl) ATCC: 2091 Lot: 04205401 (productivity) Incubate: (36 ± 1) °C για (24 ± 2) h	PR ≥ 0.7 Colonies Green	Colonies Green 442 cfu
<i>Escherichia coli</i> (≥500 cfu/10μl) ATCC 25922 Lot: 00207501 (selectivity) Incubate: (36 ± 1) °C για (24 ± 2) h	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / EAOT EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.

**Release of the batch**YES  NO **DATE**

05/04/2024

**Iniotaki Fofi****Quality Control Manager**