

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	GLUCOSE BROTH
Code	070045
Lot number	2791u128
Production date	28/03/24
Expiration date	28/03/25

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	7.1 +/- 0,2	7.0 (Confirmed)
Quantity of material	5ml (070045)	Confirmed
Color & cleanliness of the material	Clear beige broth	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
3/070045	36 ± 1 °C for 48 hours	3/070045	0


FINAL PRODUCT PERFORMANCE TEST (Incubate at 37 ± 1 °C for 18 - 24 h)

Microorganism	Expected results	Observed results
<i>Staphylococcus aureus</i> (≥500 cfu/10μl) e10.24 ATCC 29213. Lot: 01403703 (productivity)	turbidity GLU.B. (1–2)f	Confirmed
<i>Escherichia coli</i> (≥500 cfu/10μl) e10.24 ATCC 25922. Lot: 00207501. (Productivity)	turbidity GLU. B. (1–2)f	Confirmed

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- f = Development/turbidity: 0: Growth zero. 1: Weak growth. 2: Good growth

Release of the batchYES NO **DATE**

01/04/24

Iniotaki Fofi**Quality Control Manager**