

G. PAPANIKOLAOU & SIA E.E.

Production laboratories of Culture Media Potamou 5 VIOPA Kerateas PC 19001 – Greece.

Tel: +30 2299 0 66113 Fax: +30 2299 0 66112 E-mail: bioprep1@otenet.gr www.bioprepare.gr

QUALITY CONTROL CERTIFICATE

FINAL PRODUCT DATA

Description	BRILLIANT GREEN AGAR (ISO 6579)
Code	010016
Lot number	2746e23
Production date	23/01/24
Expiration date	23/04/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	6.9 +/- 0,2	6.9 (Confirmed)
Quantity of material	20ml (010016)	Confirmed
Color & cleanliness of the material	Tile clear	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products	Incubation	Number of suitable products	Number of inappropriate products
(3‰)			
2/010016	36 ± 1 °C for 48 hours	2/010016	0

FINAL PRODUCT PERFORMANCE TEST (According to ISO 6579) Incubation: (37 ± 1) °C for (24 ± 3) h

Microorganism	Expected results	Observed results
Salmonella typhimurium (≥500 cfu)	PR ≥ 0,5	(Pink – white colonies with red band)
ATCC 14028. Lot: 01102701 (Productivity)	(Pink – white colonies with red band)	320 cfu
Enterococcus faecalis (≥500 cfu)	Not growing (0)	0 cfu
ATCC 29212. Lot: 01805401 (Selectivity)		

- Production and quality control tests were carried out at the premises of Bioprepare G. Papanikolaou & Co. According to EN ISO 9001:2015 / EAOT EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.

Release of the batch	DATE	Aspasia Papanikolaou
YES V NO	28/01/24	- June

Quality Control Responsible