

G. PAPANIKOLAOU & SIA E.E.

Production laboratories of Culture Media Potamou 5 VIOPA Kerateas PC 19001 – Greece.

Tel: +30 2299 0 66113 Fax: +30 2299 0 66112 E-mail: bioprep1@otenet.gr www.bioprepare.gr

QUALITY CONTROL CERTIFICATE

FINAL PRODUCT DATA

Description	X.L.D. AGAR (ISO 6579)
Code	010130
Lot number	2738n111
Production date	11/01/24
Expiration date	11/04/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results				
Final pH at 25°C	7,4 +/- 0,2	7,4 (Confirmed)				
Quantity of material	20ml (010130)	Confirmed				
Color & cleanliness of the material	Agar red clear	Confirmed				

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
2/010130	36 ± 1 °C for 48 hours	2/010130	0

FINAL PRODUCT PERFORMANCE TEST (According to ISO 6579)

Microorganism	Expected results	Observed results
Salmonella typhimurium (≥400 cfu)	PR ≥ 0,5	(Red colonies with black center)
ATCC 14028. Lot: 01102701 (Productivity)	(Red colonies with black center)	242 cfu
Enterococcus faecalis (≥500 cfu)	Inhibited (0)	0 cfu
ATCC 29212. Lot: 01805401 (Selectivity)		

- Production and quality control tests were carried out at the premises of Bioprepare G. Papanikolaou & Co. According to EN ISO 9001: 2008. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 ° C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- PR: Productivity Rate. IV: Inoculum Value
- 0: Growth zero. 1: Weak growth. 2: Good growth

Release of the batch		DATE	Iniotaki Fofi
YES V	NO	16/01/2024	Afrika
			Quality Control Manager