

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	m – ENTEROCOCCUS AGAR (Slanetz-Bartley Medium) ISO 7899-2
Code	050063
Lot number	2734e05
Production date	05/01/24
Expiration date	05/04/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	7,2 +/- 0,2	7,2 (Confirmed)
Quantity of material	9ml (050063)	Confirmed
Color & cleanliness of the material	Beige agar light clear.	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (‰)	Incubation	Number of suitable products	Number of inappropriate products
2/050063	36 ± 1 °C for 48 hours	2/050063	0

FINAL PRODUCT PERFORMANCE TEST Incubate: (36 ± 1) °C for (24 ± 2) h

Microorganism	Expected results	Observed results
<i>Enterococcus faecalis</i> (≥300 cfu/ml) ATCC: 29212. Lot: 01805401 (Productivity)	PR ≥ 0,5 Red-brown-pink colonies	Red-brown-pink colonies 163 cfu
<i>Escherichia coli</i> (≥500 cfu) ATCC 25922. Lot: 00207501 (selectivity)	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth


Release of the batch

YES NO

DATE

10/01/2024

Iniotaki Fofi



Quality Control Manager