

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	THAYER MARTIN AGAR
Code	050111
Lot number	2749c126
Production date	26/01/24
Expiration date	26/04/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	7,3 +/- 0,2	7.3 (Confirmed)
Quantity of material	9ml (050111)	Confirmed
Color & cleanliness of the material	Coffee agar - unclear chocolate.	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3%)	Incubation	Number of suitable products	Number of inappropriate products
1/050111	36 ± 1 °C for 48 hours	1/050111	0


**FINAL PRODUCT PERFORMANCE TEST (Incubate at 37 ± 1 °C for 24 - 48 h with 10% CO<sub>2</sub>.)**

Microorganism	Expected results	Observed results
<i>Neisseria gonorrhoeae</i> (≥500 cfu) ATCC: 19424. Lot: 00502402 (productivity)	PR ≥ 0,5 Colonies small, gray, white	Colonies small, gray, white 291 cfu
<i>Escherichia coli</i> (≥500 cfu) ATCC 25922. Lot: 00207501 (selectivity)	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.

**Release of the batch**YES  NO **DATE**

29/01/2024

**Iniotaki Fofi****Quality Control Manager**