

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	Mueller Kauffmann Broth W/ Brilliant Green & Novobiocin (ISO 6579)
Code	070526
Lot number	2743u118
Production date	18/01/24
Expiration date	18/05/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	8,2 +/- 0,2	8,1 (Confirmed)
Quantity of material	10ml (070526)	Confirmed
Color & cleanliness of the material	Green with white sediment on standing	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
1/070526	36 ± 1 °C for 48 hours	1/070526	0

FINAL PRODUCT PERFORMANCE TEST Incubation: (According to ISO 6579). Incubation: (37 ± 1) °C for (24 ± 3) h

Microorganism	Expected results	Observed results
<i>Salmonella typhimurium</i> (≥500 cfu) ATCC: 14028. Lot: 01102701 (productivity)	Growth on XLD >10 colonies (black center)	Confirmed IV = 103
<i>Escherichia coli</i> (≥500 cfu) ATCC: 25922. Lot: 00207501 (selectivity)	Growth in TSA <100 colonies	Confirmed IV = 15

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. IV: Inoculum Value


Release of the batch

YES NO

DATE

23/01/2024

Iniotaki Fofi



Quality Control Manager